

Notification of a Body in the framework of a technical harmonization directive

From: Office for Registration of Medicinal Products, Medical Devices and Biocidal Products
Al. Jerozolimskie 181 C
02-222 Warsaw
Poland

To: European Commission
GROWTH Directorate-General
200 Rue de la Loi,
B-1049 Brussels.

Other Member States

Reference:

Legislation: Regulation (EU) 2017/745 on medical devices

Body name, address, telephone, fax, email, website :

TUV NORD Polska Sp. z o.o
ul. Mickiewicza 29
40-085 Katowice
Poland
+48 32 7864646
+48 32 7864601
biuro@tuv-nord.pl
www.tuv-nord.pl

Body info:

NB 2274

Tasks performed by the Body:

Last approval date: 2024-02-19

Product	Procedures	Articles /Annexes	Conditions
-I. CODES REFLECTING THE DESIGN AND INTENDED PURPOSE OF THE DEVICE -A. Active devices -2. Active non-implantable devices for imaging, monitoring and/or diagnosis -MDA 0202 Active non-implantable imaging devices utilising non-ionizing radiation	Conformity assessment based on a quality management system Conformity assessment based on assessment of technical documentation Conformity assessment based on product quality assurance	Annex IX(I) Annex IX(II) Annex XI(A)	Excluding MRI
-I. CODES REFLECTING THE DESIGN AND INTENDED PURPOSE OF THE DEVICE -A. Active devices -2. Active non-implantable devices for imaging, monitoring and/or diagnosis -MDA 0203 Active non-implantable devices for monitoring of vital physiological parameters	Conformity assessment based on a quality management system Conformity assessment based on assessment of technical documentation Conformity assessment based on product quality assurance	Annex IX(I) Annex IX(II) Annex XI(A)	
-I. CODES REFLECTING THE DESIGN AND INTENDED PURPOSE OF THE DEVICE -A. Active devices -2. Active non-implantable devices for imaging, monitoring and/or diagnosis -MDA 0204 Other active non-implantable devices for monitoring and/or diagnosis	Conformity assessment based on a quality management system Conformity assessment based on assessment of technical documentation Conformity assessment based on product quality assurance	Annex IX(I) Annex IX(II) Annex XI(A)	Excluding products used in ophthalmology
-I. CODES REFLECTING THE DESIGN AND INTENDED PURPOSE OF THE DEVICE -A. Active devices -3. Active non-implantable therapeutic devices and general active non-implantable devices -MDA 0302 Active non-implantable devices utilising non-ionizing radiation	Conformity assessment based on a quality management system Conformity assessment based on assessment of technical documentation Conformity assessment based on product quality assurance	Annex IX(I) Annex IX(II) Annex XI(A)	
-I. CODES REFLECTING THE DESIGN AND INTENDED PURPOSE OF THE DEVICE -A. Active devices -3. Active non-implantable therapeutic devices and general active non-implantable devices -MDA 0305 Active non-implantable devices for stimulation or inhibition	Conformity assessment based on a quality management system Conformity assessment based on assessment of technical documentation Conformity assessment based on product quality assurance	Annex IX(I) Annex IX(II) Annex XI(A)	

Product	Procedures	Articles /Annexes	Conditions
-I. CODES REFLECTING THE DESIGN AND INTENDED PURPOSE OF THE DEVICE -A. Active devices -3. Active non-implantable therapeutic devices and general active non-implantable devices -MDA 0306 Active non-implantable devices for extra-corporal circulation, administration or removal of substances and haemapheresis	Conformity assessment based on a quality management system Conformity assessment based on assessment of technical documentation Conformity assessment based on product quality assurance	Annex IX(I) Annex IX(II) Annex XI(A)	Including only infusion pumps, devices for dialysis, anaesthesia machines and devices for administration or removal of substances
-I. CODES REFLECTING THE DESIGN AND INTENDED PURPOSE OF THE DEVICE -A. Active devices -3. Active non-implantable therapeutic devices and general active non-implantable devices -MDA 0307 Active non-implantable respiratory devices	Conformity assessment based on a quality management system Conformity assessment based on assessment of technical documentation Conformity assessment based on product quality assurance	Annex IX(I) Annex IX(II) Annex XI(A)	Excluding hyperbaric chambers
-I. CODES REFLECTING THE DESIGN AND INTENDED PURPOSE OF THE DEVICE -A. Active devices -3. Active non-implantable therapeutic devices and general active non-implantable devices -MDA 0311 Active non-implantable dental devices	Conformity assessment based on a quality management system Conformity assessment based on assessment of technical documentation Conformity assessment based on product quality assurance	Annex IX(I) Annex IX(II) Annex XI(A)	
-I. CODES REFLECTING THE DESIGN AND INTENDED PURPOSE OF THE DEVICE -A. Active devices -3. Active non-implantable therapeutic devices and general active non-implantable devices -MDA 0312 Other active non-implantable surgical devices	Conformity assessment based on a quality management system Conformity assessment based on assessment of technical documentation Conformity assessment based on product quality assurance	Annex IX(I) Annex IX(II) Annex XI(A)	

Product	Procedures	Articles /Annexes	Conditions
-I. CODES REFLECTING THE DESIGN AND INTENDED PURPOSE OF THE DEVICE -A. Active devices -3. Active non-implantable therapeutic devices and general active non-implantable devices -MDA 0313 Active non-implantable prostheses, devices for rehabilitation and devices for patient positioning and transport	Conformity assessment based on a quality management system Conformity assessment based on assessment of technical documentation Conformity assessment based on product quality assurance	Annex IX(I) Annex IX(II) Annex XI(A)	Excluding active prostheses
-I. CODES REFLECTING THE DESIGN AND INTENDED PURPOSE OF THE DEVICE -A. Active devices -3. Active non-implantable therapeutic devices and general active non-implantable devices -MDA 0315 Software	Conformity assessment based on a quality management system Conformity assessment based on assessment of technical documentation Conformity assessment based on product quality assurance	Annex IX(I) Annex IX(II) Annex XI(A)	
-I. CODES REFLECTING THE DESIGN AND INTENDED PURPOSE OF THE DEVICE -A. Active devices -3. Active non-implantable therapeutic devices and general active non-implantable devices -MDA 0316 Medical gas supply systems and parts thereof	Conformity assessment based on a quality management system Conformity assessment based on assessment of technical documentation Conformity assessment based on product quality assurance	Annex IX(I) Annex IX(II) Annex XI(A)	
-I. CODES REFLECTING THE DESIGN AND INTENDED PURPOSE OF THE DEVICE -A. Active devices -3. Active non-implantable therapeutic devices and general active non-implantable devices -MDA 0317 Active non-implantable devices for cleaning, disinfection and sterilisation	Conformity assessment based on a quality management system Conformity assessment based on assessment of technical documentation Conformity assessment based on product quality assurance	Annex IX(I) Annex IX(II) Annex XI(A)	
-I. CODES REFLECTING THE DESIGN AND INTENDED PURPOSE OF THE DEVICE -B. Non-active devices -1. Non-active implants and long term surgically invasive devices -MDN 1102 Non-active osteo- and orthopaedic implants	Conformity assessment based on a quality management system Conformity assessment based on assessment of technical documentation Conformity assessment based on product quality assurance	Annex IX(I) Annex IX(II) Annex XI(A)	Excluding bone graft substitute for orthopaedic indications, knee, shoulder and hip joint replacement, hyaluronic acid implant for intra-articular use, bone cement

Product	Procedures	Articles /Annexes	Conditions
-I. CODES REFLECTING THE DESIGN AND INTENDED PURPOSE OF THE DEVICE -B. Non-active devices -1. Non-active implants and long term surgically invasive devices -MDN 1103 Non-active dental implants and dental materials	Conformity assessment based on a quality management system Conformity assessment based on assessment of technical documentation Conformity assessment based on product quality assurance	Annex IX(I) Annex IX(II) Annex XI(A)	
-I. CODES REFLECTING THE DESIGN AND INTENDED PURPOSE OF THE DEVICE -B. Non-active devices -1. Non-active implants and long term surgically invasive devices -MDN 1104 Non-active soft tissue and other implants	Conformity assessment based on a quality management system Conformity assessment based on assessment of technical documentation Conformity assessment based on product quality assurance	Annex IX(I) Annex IX(II) Annex XI(A)	Including only urological tapes, surgical meshes ligament and tendon prostheses made of multifilament polyester fibers
-I. CODES REFLECTING THE DESIGN AND INTENDED PURPOSE OF THE DEVICE -B. Non-active devices -2. Non-active non-implantable devices -MDN 1202 Non-active non-implantable devices for administration, channelling and removal of substances, including devices for dialysis	Conformity assessment based on a quality management system Conformity assessment based on assessment of technical documentation Conformity assessment based on product quality assurance	Annex IX(I) Annex IX(II) Annex XI(A)	
-I. CODES REFLECTING THE DESIGN AND INTENDED PURPOSE OF THE DEVICE -B. Non-active devices -2. Non-active non-implantable devices -MDN 1204 Non-active non-implantable devices for wound and skin care	Conformity assessment based on a quality management system Conformity assessment based on assessment of technical documentation Conformity assessment based on product quality assurance	Annex IX(I) Annex IX(II) Annex XI(A)	Excluding sutures
-I. CODES REFLECTING THE DESIGN AND INTENDED PURPOSE OF THE DEVICE -B. Non-active devices -2. Non-active non-implantable devices -MDN 1208 Non-active non-implantable instruments	Conformity assessment based on a quality management system Conformity assessment based on assessment of technical documentation Conformity assessment based on product quality assurance	Annex IX(I) Annex IX(II) Annex XI(A)	
-I. CODES REFLECTING THE DESIGN AND INTENDED PURPOSE OF THE DEVICE -B. Non-active devices -2. Non-active non-implantable devices -MDN 1209 Non-active non-implantable dental materials	Conformity assessment based on a quality management system Conformity assessment based on assessment of technical documentation Conformity assessment based on product quality assurance	Annex IX(I) Annex IX(II) Annex XI(A)	

Product	Procedures	Articles /Annexes	Conditions
-I. CODES REFLECTING THE DESIGN AND INTENDED PURPOSE OF THE DEVICE -B. Non-active devices -2. Non-active non-implantable devices -MDN 1211 Non-active non-implantable devices for disinfecting, cleaning and rinsing	Conformity assessment based on a quality management system Conformity assessment based on assessment of technical documentation Conformity assessment based on product quality assurance	Annex IX(I) Annex IX(II) Annex XI(A)	
-I. CODES REFLECTING THE DESIGN AND INTENDED PURPOSE OF THE DEVICE -B. Non-active devices -2. Non-active non-implantable devices -MDN 1212 Non-active non-implantable devices for processing and preservation of human cells, tissue or organs including in vitro fertilisation (IVF) and assisted reproductive technologies (ART)	Conformity assessment based on a quality management system Conformity assessment based on assessment of technical documentation Conformity assessment based on product quality assurance	Annex IX(I) Annex IX(II) Annex XI(A)	Excluding devices for in vitro fertilisation (IVF) and assisted reproductive technologies (ART)
-I. CODES REFLECTING THE DESIGN AND INTENDED PURPOSE OF THE DEVICE -B. Non-active devices -2. Non-active non-implantable devices -MDN 1213 Non-active non-implantable devices composed of substances to be introduced into the human body via a body orifice or the dermal route	Conformity assessment based on a quality management system Conformity assessment based on assessment of technical documentation Conformity assessment based on product quality assurance	Annex IX(I) Annex IX(II) Annex XI(A)	
-I. CODES REFLECTING THE DESIGN AND INTENDED PURPOSE OF THE DEVICE -B. Non-active devices -2. Non-active non-implantable devices -MDN 1214 General non-active non-implantable devices used in health care and other non-active non-implantable devices	Conformity assessment based on a quality management system Conformity assessment based on assessment of technical documentation Conformity assessment based on product quality assurance	Annex IX(I) Annex IX(II) Annex XI(A)	Including only ultrasound gels, medication cups

Horizontal technical competences	Limitations
MDS 1001 Devices incorporating medicinal substances:	

Horizontal technical competences	Limitations
MDS 1005 Devices in sterile condition:	Including: aseptic processing, ethylene oxide gas sterilisation (EOG), moist heat sterilisation, radiation sterilisation, filtration
MDS 1006 Reusable surgical instruments:	
MDS 1007 Devices incorporating or consisting of nanomaterial:	
MDS 1008 Devices utilising biologically active coatings and/or materials or being wholly or mainly absorbed or locally dispersed in the human body or are intended to undergo a chemical change in the body:	
MDS 1009 Devices incorporating software/utilising software/controlled by software, including devices intended for controlling, monitoring or directly influencing the performance of active or active implantable devices:	
MDS 1010 Devices with a measuring function:	
MDS 1011 Devices in systems or procedure packs:	
MDS 1012 Products without an intended medical purpose listed in Annex XVI to Regulation (EU) 2017/745:	Limited to devices emitting electromagnetic radiation for use on the human body according to Annex XVI p. 5
MDS 1013 Class III custom-made implantable devices:	
MDT 2001 Devices manufactured using metal processing:	
MDT 2002 Devices manufactured using plastic processing:	
MDT 2003 Devices manufactured using non-metal mineral processing (e.g. glass, ceramics):	Excluding processing of glass
MDT 2004 Devices manufactured using non-metal non-mineral processing (e.g. textiles, rubber, leather, paper):	Excluding processing of leather
MDT 2005 Devices manufactured using biotechnology:	
MDT 2006 Devices manufactured using chemical processing:	
MDT 2007 Devices which require knowledge regarding the production of pharmaceuticals:	
MDT 2008 Devices manufactured in clean rooms and associated controlled environments:	
MDT 2009 Devices manufactured using processing of materials of human, animal, or microbial origin:	Excluding processing of animal materials
MDT 2010 Devices manufactured using electronic components including communication devices:	

Horizontal technical competences	Limitations
MDT 2011 Devices which require packaging, including labelling:	
MDT 2012 Devices which require installation, refurbishment:	Limited to electronic devices and medical gas installations